

IN THE UNITED STATES PATENT AND TRADEMARK OFFICE

Applicant: MURPHY, Kieran

Examiner: BUI, Vy Q.

Serial No.: 10/727,667

Art Unit: 3773

Filed: December 5, 2003

Docket No.: P1677US03 (8131-41)

For: **DEVICE VIEWABLE UNDER AN IMAGING BEAM**

STATEMENT OF SUBSTANCE OF INTERVIEW

Commissioner for Patents

P.O. Box 1450

Alexandria, VA 22313-1450

Commissioner:

Pursuant to the Examiner Interview conducted on November 23, 2010, please enter in above-identified application the statement which begins on page 2 of this paper.

STATEMENT OF SUBSTANCE OF INTERVIEW

On November 23, 2010 the undersigned conducted a telephone interview with the Examiner. No exhibit was shown nor was a demonstration conducted.

The discussion focused on the cited Fearnot reference wherein both the Examiner and the undersigned concluded that the Fearnot reference does not disclose a stent having an image density of less than about 1200 Hounsfield Units as recited in dependent Claim 25.


The Examiner agreed to amend Claim 1 to include the allowable subject matter of Claim 25, whereupon Claim 25 would be canceled, and issue a Notice of Allowance.

It was also agreed that Withdrawn Claims 2, 3, 5 – 9 and 17 – 23 would be canceled.

Respectfully submitted,

F. CHAU & ASSOCIATES, LLC

Date: DECEMBER 28, 2010

By 
Richard J. Paciulan
Reg. No. 28,248

Mailing Address:

F. CHAU & ASSOCIATES, LLC

130 Woodbury Road

Woodbury, New York 11797

TEL: (516) 692-8888

FAX: (516) 692-8889